

CALIBRATION CERTIFICATE

CUSTOMER
CROSSWAYS RECYCLING PLOT 15, MANOR WAY BUSINESS PARK MANOR WAY SWANSCOMBE KENT DA10 0PP

CERTIFICATE NO: CRL.1117.3	
TEST DATE 18/10/2019	NEXT TEST OCTOBER 2020

WEIGHT INFORMATION		
WEIGHT CERT No TR10663	CLASS M1	DATE TESTED 20/03/2019
WEIGHT SET No 1-23A		CERT INTERVAL 12M

EQUIPMENT DETAILS			
MAKE: IWS - DINI ARGEO	CAPACITY: 50000 kg	VERIFIED: YES	
MODEL: 153/PM6/DFWL	INCREMENTS: 20 kg	CUSTOMER REF: WEIGHBRIDGE	
SERIAL No: 0100594170	WORKING CAP: 44000 kg	LOCATION: OFFICE / YARD	

LINEARITY AND HYSTERESIS TEST					
AS FOUND TEST			DEFINITIVE TEST		
LOAD kg	READING kg	ERROR kg	LOAD kg	READING kg	ERROR kg
0	0	0	0	0	0
1000	1000	0	1000	1000	0
10000	10000	0	10000	10000	0
20000	20000	0	20000	20000	0
29520	29520	0	29520	29520	0
39520	39500	-20	39520	39500	-20
43400	43360	-40	43400	43360	-40

ECCENTRIC TEST			
AS FOUND			
POSITION	LOAD kg	READING kg	ERROR kg
1	5880	5880	0
2	5880	5860	-20
3	5880	5880	0
4	5880	5880	0
5	5880	5860	-20
6	5880	5880	0
7			
8			

ECCENTRIC TEST			
DEFINITIVE			
POSITION	LOAD kg	READING kg	ERROR kg
1	5880	5880	0
2	5880	5860	-20
3	5880	5880	0
4	5880	5880	0
5	5880	5860	-20
6	5880	5880	0
7			
8			

SKETCH							
<table border="1"> <tr> <td>2</td> <td>4</td> <td>6</td> </tr> <tr> <td>1</td> <td>3</td> <td>5</td> </tr> </table>	2	4	6	1	3	5	INDICATOR
2	4	6					
1	3	5					

REPEATABILITY TEST				
LOAD kg	1ST kg	2ND kg	3RD kg	DEVIATION
43360	43360	43360	43360	0 kg

ENGINEERS NOTES
WEIGHBRIDGE TESTED AND CALIBRATED AS NECESSARY

TOLERANCES: +/-		Certificate 1 of 1
RANGE	TOLERANCE	
0 - 10000 kg	20 kg	
10000 kg - 40000 kg	40 kg	
40000 kg - 50000 kg	60 kg	