

**CALIBRATION CERTIFICATE**

<b>CUSTOMER</b>
CROSSWAYS RECYCLING PLOT 15, MANOR WAY BUSINESS PARK MANOR WAY SWANSCOMBE KENT DA10 0PP

<b>CERTIFICATE NO: CRL1020.1</b>	
TEST DATE 29/10/2020	NEXT TEST OCTOBER 2021

<b>WEIGHT INFORMATION</b>		
WEIGHT CERT No TR10945	CLASS M1	DATE TESTED 20/03/2020
WEIGHT SET No 1-23A		CERT INTERVAL 12M

**EQUIPMENT DETAILS**

MAKE: IWS - DINI ARGEO	CAPACITY: 50000 kg	VERIFIED: YES
MODEL: 153/PM6/DFWL	INCREMENTS: 20 kg	CUSTOMER REF: WEIGHBRIDGE
SERIAL No: 0100594170	WORKING CAP: 44000 kg	LOCATION: OFFICE / YARD

**LINEARITY AND HYSTERESIS TEST**

AS FOUND TEST		
LOAD kg	READING kg	ERROR kg
0	0	0
1000	1000	0
10000	10000	0
20000	20000	0
30020	30000	-20
40020	40000	-20
44480	44440	-40

DEFINITIVE TEST		
LOAD kg	READING kg	ERROR kg
0	0	0
1000	1000	0
10000	10000	0
20000	20000	0
30020	30020	0
40020	40000	-20
44480	44460	-20

**ECCENTRIC TEST**

AS FOUND			
POSITION	LOAD kg	READING kg	ERROR kg
1	6460	6460	0
2	6460	6460	0
3	6460	6440	-20
4	6460	6440	-20
5	6460	6440	-20
6	6460	6440	-20
7			
8			

**ECCENTRIC TEST**

DEFINITIVE			
POSITION	LOAD kg	READING kg	ERROR kg
1	6460	6460	0
2	6460	6460	0
3	6460	6460	0
4	6460	6460	0
5	6460	6440	-20
6	6460	6440	-20
7			
8			

**SKETCH**

2	4	6
1	3	5

INDICATOR

**REPEATABILITY TEST**

LOAD kg	1ST kg	2ND kg	3RD kg	DEVIATION
44460	44460	44460	44440	20 kg

**ENGINEERS NOTES**

WEIGHBRIDGE TESTED AND CALIBRATED AS NECESSARY
---

**TOLERANCES: +/-**

RANGE	TOLERANCE
0 - 10000 kg	20 kg
10000 kg - 40000 kg	40 kg
40000 kg - 50000 kg	60 kg

Certificate  
1 of 1